

Declaration of Conformity

We Techno Medica Co., Ltd.
5-5-1, Nakamachidai, Tsuzuki-ku, Yokohama 224-0041 JAPAN
declare in our sole responsibility that the following product conforms to all the relevant provisions.

- Type of Equipment : Clinical Analyzer for POC testing
- Model Name : GASTAT-mini
Sensor card 981, Sensor card 982,
Sensor card 983, Sensor card 984, Sensor card 985
- Council Directives : IVDD 98/79/EC
- Standards(s) to which Conformity declared :
EN61010-1: 1993+A2:1995
- Year of Manufacture : 1999

Signature: Hiroki Yamazaki
Name: Hiroki Yamazaki
Title: Senior Manager, Product development div.
Date: Jan 6, 2003

Declaration of Conformity

We Techno Medica Co., Ltd.
5-5-1, Nakamachidai, Tsuzuki-ku, Yokohama 224-0041 JAPAN
declare in our sole responsibility that the following product conforms to all the relevant provisions.

- Type of Equipment : Clinical Analyzer for POC testing
- Model Name : GASTAT-mini
Sensor card 981, Sensor card 982,
Sensor card 983, Sensor card 984, Sensor card 985
- Council Directives : Low Voltage Directive, 73/23/EEC, 93/68/EEC
- Standards(s) to which Conformity declared :
EN61010-1: 1993+A2:1995
- Year of Manufacture : 1999

Signature: Hiroki Yamazaki
Name Hiroki Yamazaki
Title Senior Manager, Product development div.
Date Oct 15, 1999

Declaration of Conformity

We Techno Medica Co., Ltd.

5-5-1, Nakamachidai, Tsuzuki-ku, Yokohama 224-0041 JAPAN
declare in our sole responsibility that the following product conforms
to all the relevant provisions.

- Product Name: Blood Gas System GASTAT-600 Series
- Model Name : GASTAT600/601/602i/603ie
- Council Directives :
 Low Voltage Directive 73/23/EEC Amended Directive.
 89/336/EEC EN61326:1997/A1:1998/A2:2001
- Standards(s) to which Conformity declared :
 EN61010-1: 2001(2nd Edition)
 EN61326:1997/A1:1998/A2:2001

Signature: *Hiroki Yamazaki*

Name Hiroki Yamazaki

Title Senior Manager, Product development div.

Date Dec 1, 2003

Declaration of Conformity

We Techno Medica Co., Ltd.

5-5-1, Nakamachidai, Tsuzuki-ku, Yokohama 224-0041 JAPAN

declare in our sole responsibility that the following product conforms to all the relevant provisions.

- Product Name: Blood Gas System GASTAT-600 Series
- Model Name : GASTAT600/601/602i/603ie
- Council Directives :
IVDD 98/79/EC
89/336/EEC EN61326:1997/A1:1998/A2:2001
- Standards(s) to which Conformity declared :
EN61010-1: 2001(2nd Edition)
EN61326:1997/A1:1998/A2:2001

Signature: Hiroki Yamazaki

Name Hiroki Yamazaki

Title Senior Manager, Product development div.

Date Dec 1, 2003