

WITH EUROPEAN DIRECTIVE

Order No. 20219

A sample of the following product has been tested and is stated by Nemko to be in conformity with the applicable European safety- and EMC standards referred below.

Manufacturer APEL Co., Ltd.,
1155-4, Angyouryou Negishi,
Kawaguchi-shi, Saitama 333-0834, Japan

Product Photoelectric Colorimeter

Model/type AP-101

Data 10W 220V AC 50Hz

Other specification Cl. I

Standards applied Safety std.: EN 61010-1: 2nd. ed. 2001
EMC std.: EN 55022:1998 + A1:2000 + A2:2003
EN 61000-3-2:2000
EN 61000-3-3_1995 + A1:2001
EN 55024:1998 + A1:2001 + A2:2003


Statement reference 20219

It may therefore be presumed that the tested sample of the product is in conformity with the technical provisions of the following European Directives including the latest amendments, and with national legislation implementing these Directives:

- Low Voltage Directive 73/23/EEC, including 93/68/EEC
- EMC Directive 89/336/EEC, including 92/31/EEC and 93/68/EEC

On this basis, the manufacturer (or the European authorized representative), may draw up an EC/EEA Declaration of Conformity and affix the CE-marking as indicated below to each conforming product.

Date of issue 04 June 2004



signature

Knut Mathisen

Certification Department



Nemko AS
P.O. Box 73, Blindern
N-0314 Oslo, Norway

Office address
Gaustadalléen 30
Oslo

Telephone
+47 22 96 03 30
Enterprise number:

Fax
+47 22 96 05 50
NO 974404532

LETTER OF COMPLIANCE

Ref. No. 199917211

Product	Spectrophotometer
Name and address of the applicant	Zintex Co., Ltd., Kudan Minami 3-9-11-1206, Chiyoda-Ku, Tokyo 102-0074 Japan
Name and address of the manufacturer	Apel Co. Ltd., 1155-4, Angyouryou Negishi, Kawaguchi-Shi, Saitama 333-0834, Japan
Name and address of the factory	Apel Co. Ltd., 1155-4, Angyouryou Negishi, Kawaguchi-Shi, Saitama 333-0834, Japan
Rating and principal characteristics	15W 230V 50/60Hz, Cl. I
Trade mark (if any).	Apel
Model/Type Ref.	PD-303
Additional information (if necessary)	
A sample of the product has been tested and found to be in conformity with the current EN and equivalent standard, (number and edition).	EN 61010-1 (1993) + A2 (1995) IEC 1010-1 (1990) + A1 (1992) + A2 (1995)
EMC/radio interference measurement	Nemko test report 1999 19209

This "Letter of compliance" is the result of testing a sample of the product submitted, in accordance with the provisions of the relevant specific standard.

This "Letter of compliance" is not a certificate neither must be used as one.

Oslo, date 2000-03-13

Issued by

**Nemko**Arild Hansgaard
Principal Engineer
Head of SectionInternal ref. **Vegard Andersen**
(No., Name or initials of contact persons)

**WITH EUROPEAN DIRECTIVE**Order No. **2001 1117**

A sample of the following product has been tested and is stated by NEMKO to be in conformity with the applicable European safety- and EMC standards referred below.

Manufacturer	Apel Co. Ltd., 1155-4, Angyouryou Negishi, Kawaguchi-Shi, Saitama 333-0834, Japan		
Product	Dual wavelength total bilirubin meter		
Group 61 69000	Various equipment for laboratory use		
Model/type	BR 501		
Data	35W 230VAC 50/60Hz		
Other specification	Cl. I product		
Standards applied	Safety: EN 61010-1 (1993) + A2 (1995) IEC 61010-1 (1990) + A1 (1992) + A2 (1995) EMC: EN 61326-1 (1997) + A1 (1998)		
Statement reference	Safety:	Nemko ref.	2001 11177
	EMC:	Nemko ref.	2001 13176

It may therefore be presumed that the tested sample of the product is in conformity with the technical provisions of the following European Directives including the latest amendments, and with national legislation implementing these Directives:

- Low Voltage Directive 73/23/EEC
- EMC Directive 89/336/EEC

On this basis, the manufacturer (or the European authorized representative), may draw up an EC/EEA Declaration of Conformity and affix the CE-marking as indicated below to each conforming product.

Remarks Confer the attached covering letter

Date of issue 4 April 2001



signature

Arild Hansgaard
Head of section



Nemko AS
P.O. Box 73, Blindern
N-0314 Oslo, Norway

Office address
Gaustadalléen 30
Oslo

Telephone
+47 22 96 03 30
Enterprise number:

Fax
+47 22 96 05 50
NO 974404532

WITH EUROPEAN DIRECTIVE

Order No. 2000 26269

A sample of the following product has been tested and is stated by NEMKO to be in conformity with the applicable European safety- and EMC standards referred below.

Manufacturer	Apel Co. Ltd., 1155-4, Angyouyou Negishi, Kawaguchi-Shi, Saitama 333-0834, Japan
Product	Spectrophotometer
Group 61 69000	Various equipment for laboratory use.
Model/type	PD-303s
Data	9W 230VAC 50/60Hz
Other specification	Cl. I product
Standards applied	Safety: EN 61010-1 (1993) + A2 (1995) IEC 1010-1 (1990) + A1 (1992) + A2 (1995) EMC: EN 61326-1 (1997) + A1 (1998)
Statement reference	Safety : Nemko ref. 2000 26269 EMC : Nemko ref. 1999 19209

It may therefore be presumed that the tested sample of the product is in conformity with the technical provisions of the following European Directives including the latest amendments, and with national legislation implementing these Directives:

- Low Voltage Directive 73/23/EEC
- EMC Directive 89/336/EEC

On this basis, the manufacturer (or the European authorized representative), may draw up an EC/EEA Declaration of Conformity and affix the CE-marking as indicated below to each conforming product.

Remarks Confer the attached covering letter

Date of issue 31 July 2000

Frank Skjerve
signature

pv Arild Hansgaard
Head of section

